

PS01.04.10 TIME-RESOLVED PROTEIN DATA COLLECTION SYSTEM WITH LARGE IMAGING PLATE. K.Sakabe, N.Kamiya¹, N. Watanabe², S. Adachi¹, K. Sasaki³, S.Ikemizu⁴, T.Higashi⁵, & N. Sakabe⁴. Dept. of Chem., Nagoya Univ., Chikusa, Nagoya, 464 Japan, ¹RIKEN, Hirosawa 2- 1, Wako, Saitama, 351 01 Japan, ²PF, KEK, Tsukuba 305 Japan, ³College of Medical Technology, Nagoya Univ. Higashi, Nagoya, 461, ⁴Institute of Applied Biochem., Univ. of Tsukuba, Tsukuba, Ibaraki 305 Japan, ⁵Rigaku Corporation Matsubara, Akishima, Tokyo 196 Japan

We have developed a data collection system which can fit for both time resolved Laue and LOT, and name it as time-resolved camera system. Typical nature of this camera for time-resolved Laue is that imaging plate cassette with 800x800mm detector area can be moved quickly while rotating fast shutter is chopping the X-ray to get m-sec order of time resolution spots whose images are aligned along horizontal direction. Chopping is better than the streaks as following reasons; 1. Easy to get accurate integrate intensity data. 2. Back ground can be reduced extensively. 3 To reduce the dose of X-ray to the crystal and to reduce the crystal damage. The evaluation of this system has been done by crystal of ω -amino acid aminotransferase whose space group I222 and cell dimensions are $a=137.9$, $b=124.7$, and $c=61.5\text{\AA}$. The normal Laue data collected from 42 frames with 400x800mm. The recovery is 62% within 2\AA resolution. The R (I) is 8.4%. Time resolved Laue data was collected from three shots at 2mm interval. The R factor(I) between three spots is 0.07 for 3580 reflection in a frame which corresponds to 75% of single spots exposure in the same condition. Time-resolved experiment using LOT has been done by this system switching to as the Weissenberg camera In the case of tetragonal lysozyme using flow cell, independent data up to 1.8\AA resolution has been collected within 15 min with two frames. Rmerge(I) is 0.045.

PS01.04.11 ROTATING ANODE/AREA DETECTOR DATA COLLECTION ON AXES >300 Å AND NEW, RAPID CALIBRATION METHOD. James C. Phillips, Siemens Energy and Automation, Inc. Analytical Instrumentation, 6300 Enterprise Lane, Madison, WI 53719-1173

Crystallographers are investigating ever larger unit cells. HI-STAR with high resolution mode and Dual HI-STAR systems were designed to meet this challenge. The Siemens HI-STAR multiwire area detector has previously been calibrated with an Fe^{55} radioactive source, as have the X100 and the X1000, earlier models. This procedure has taken several hours when the system is configured for large unit cells. However, the new calibration method, using amorphous Iron foil placed at the crystal position and generating fluorescence by irradiating it with $\text{CuK}\alpha$ radiation is much faster. For example, using a dual HI-STAR at 300/452 mm from the sample, on a rotating anode generator calibrations were done for both detectors simultaneously for flood field (15 minutes) and spatial (15 minutes). The 300 mm detector was in 1024x1024 pixel mode, the 452 mm detector was in 512x152 mode. An intensity comparison was made with a 2.92 year old 100 μC Fe^{55} source (reduced to 46% of original activity) for the dual detector configuration above. Foil count rates were 43 times stronger for the 300 mm, 46 times for the 452 mm. With calibration by this method data was collected on a crystal with a 346\AA axis and one 400\AA . The cell parameters and the statistics of the data integrated by SAINT clearly show the power of the dual HI-STAR system for these "front-line" experiments and that calibrations are accurate as well as convenient.

PS01.04.12 AN ANALYSIS OF DATA COLLECTION STRATEGIES AND DATA REDUCTION SOFTWARE FOR IMAGE PLATE DATA. Bing Hu¹, John Rose² and Bi-Cheng Wang², ¹Dept. of Crystallography, University of Pittsburgh, Pittsburgh, PA 15260, U.S.A., ²Dept. of Biochemistry & Molecular Biology, University Georgia, Athens, GA 30602, U.S.A.

In designing a strategy for macromolecular data collection, one of the most often asked questions is "What scan range should I use?". The answer may vary depending on the facilities used for data collection and the programs used for data processing. Intuitively, narrow oscillation data slices (0.1 - 0.25°) should give a better signal-to-noise ratio than large oscillation data slices (say, 0.5 - 2.0°). However most image plate data is collected using a large oscillation data slice due in part to the slow readout time of the commercial detectors, the decay of diffraction quality with time and the initial lack of data reduction software for narrow oscillation data slices. In addition, there appears to be no reported systematic study on the relation of scan range to data quality for image plate data which prompted us to do a systematic study on this subject. Since data quality may also be affected by the techniques (e.g. 2D versus 3D profile fitting) used in the data reduction, we have included data quality versus data reduction software used as part of our study.

The data used in this study was a 2.6\AA set of 0.25° oscillation data collected at room temperature using a Mar Research 30 cm image plate scanner on crystals of the Neurophysin-hydrin I complex (space group $P4_12_12$, $a=b=68.7\text{\AA}$ and $c=113.64\text{\AA}$). The 0.25° images were combined to form the 0.5° , 0.75° , 1.00° , 1.25° and 1.5° used in the analysis. Each data set was then processed using X-GEN, XDS, MOSFILM and DENZO. The results of our analysis will be presented.

Work supported by NIH grant GM-41936.

Synchrotron Radiation III - Applications Time Resolved Micro-crystal High Energy

MS01.05.01 STUDIES OF THE SYNTHESIS, AND TRANSFORMATION OF MATERIALS USING IN SITU TIME-RESOLVED POWDER DIFFRACTION. J. C. Hanson, J. Aruajo, P. Norby, Brookhaven Nat. Lab., USA; A. N. Christensen, Aarhus Univ., Denmark; K. Ståhl, DTU, Lyngby, Denmark; G. Artioli, Univ. of Milan, Italy; A. Gualtieri, Univ. of Modena, Italy.

Time-resolved synchrotron powder diffraction data will be presented that have been collected in situ with position sensitive detectors in order to study the kinetics of hydrothermal syntheses¹ and phase changes².

The hydrothermal syntheses of Co and Mg substituted aluminophosphates have been found to form many different MAPO frameworks depending on the template, pH and hydrothermal conditions.

The phase transformations of KNO_3 in the range 303 to 533K have been shown to be dependent on the thermal history of the sample.

We have also found that the in-situ time-resolved data can be used for Rietveld profile refinements³. For example, the study of laumontite⁴ offers great insight on the fine details of water molecules-cation-framework oxygen atoms interaction during the thermally driven release of the water molecules from the zeolitic channels. A molecular movie showing the steps in the dehydration of laumontite is available on the WWW at <http://www.chemistry.bnl.gov/x7b/x7bhome.html>. In stilbite, the site distribution of the water molecules is more complex, but the study clearly shows the temperature-dependent shift of the cations towards the framework oxygens, finally leading to a first-order phase